

Results:

Test Article	Percent VFE (%)
KNOS-1001GM	>99.9 ^a
KNOS 1003GM	>99.9 ^a
KNOS 1005GM	>99.9 ^a

^a There were no detected plaques on any of the Andersen sampler plates for this test article.

The filtration efficiency percentages were calculated using the following equation:

$$\% VFE = \frac{C - T}{C} \times 100$$

C = Positive control average

T = Plate count total recovered downstream of the test article

Note: The plate count total is available upon request